

391U DENSITOMETER

Featured Specifications

Film Measures X-Rite 21-step sensitometer formats exposed on 8" length or longer films.

NOTE:

The exposed sensitometric film must have a sensitometric exposure with a gamma of .7 or greater on steps 7 through 15. The density must increase a minimum of .11 D between each step.



Measurement Speed 1.2" per second
Spectral Response ANSI Visual
Density Range 0 - 4.5D
Density Accuracy $\pm .02D$ (0 - 3.00D)
(ANSI PH2.19-1986) $\pm 2.0\%$ (3.01D - 4.00D)

NOTE:

When compared to other densitometers, density variations may exist due to calibration and spectral differences.

Density Repeatability $\pm .01D$ (0 - 3.00D)
 $\pm 1.0\%$ (3.01 - 3.50D)
 $\pm 2.0\%$ (3.51 - 4.00D)
Voltage Requirements 12v DC, Universal 100-240VAC Adapter;
50/60Hz P/N SE30-77
Dimensions 7.2" x 6.0" x 2.75"
182.8mm x 152.4mm x 69.8mm

Distributed by:
www.radcal.com

Radcal®

Mfg by: x-rite

426 W. Duarte Rd. • Monrovia, CA 91016 • P. 626.357-7921 • F. 626.357-8863